Handheld Multifunction Data Logger



Temperature - Relative Humidity - Atmospheric and differntial pressure - Air speed - Illuminance -Irradiance - CO₂- Direct voltage and current

○ GREAT FLEXIBILITY

Three indipendent input channels - Interchangeable probes - automatic recognition of the probes

O DATA IMMEDIATELY AVAILABLE

Simultaneous display of 3 variables in numerical form Real time visualization on display of the graph of a measured variable

○ ACCURATE AND RELIABLE

All **probes** are supplied **factory-calibrated** and do not require additional interventions of the user

○ LONG & SAFE LOGGING

Data logging function with data storage in CSV format directly into the SD card - Long logging time -Configurable storage interval & automatic switch-off



HD31 is a portable multifunctional data logger instrument with color graphic LCD display.

It is equipped with **3 independent inputs** and each input can be connected to SICRAM probes for the measurement of a **large variety of physical quantities**. The SICRAM module allows the instrument to immediatly recognize the type of probe connected. Moreover, it stores into memory the probe calibration data: in this way the probe can be indifferently connected to the three inputs of the instrument or to another instrument with no need to be recalibrated. All probes are supplied **factory-calibrated** and are interchangeable.

The instrument's inputs can also be connected to non SICRAM probes, by using the relevant interface modules between instrument and probe.

Measurement units are selectable according to the measured physical quantity. The display can show up to 36 quantities, three of which can be displayed simultaneously in numeric format. A real-time measured quantity graph is also shown in the display.

Data logging function with data storing in CSV format directly to the SD type memory card, for a **long duration of the logging**. Storage interval configurable by the user.

USB port with mini-USB connector for PC connection, for configuration and download of the acquired data.

Automatic or manual start and stop of the logging. Storing of date and time of each recorded sample. Automatic creation of measurement **reports in PDF** format on the memory card.

PRC Technologies Corp., Ltd. ลาดพร้าว 101 กรุงเทพ 10240 โทรศัพท : 02 530 1714, 02 932 1711 มือถือ : 086 360 8600 อีเมล : contact@prctech.net LINE ID1 : prctec-info, LINE ID2 : @prctec



Technical Specifications of the instrument

Power supply	Rechargeable 3.7 V Li-ion internal battery, 2250 mA/h capacity, 3-pole JST connector. External power supply unit (SWD05), to be connected to the mini-USB connector of the instrument. If connected to PC, it is powered from the computer's USB port (500 mA at least).						
Battery autonomy	18 hours of continuous operation (typ. autonomy with full charge battery and three connected Pt100 probes). The effective autonomy depends on the number and type of connected sensors.						
Logging interval	1, 5, 10, 15, 30 s 1, 2, 5, 10, 15, 20, 30 m / 1 hour						
Storage capacity	SD memory card with capacity up to 8 GB						
Inputs	3 inputs 8-pole DIN45326 connector						
Accuracy @ 20°C	± 0.02 % of the measure (excluding probes error)						
Temperature drift @ 20°C	20 ppm/°C (excluding probes drift)						
Long term stability	0.05 %/year (excluding probes stability)						
Clock stability	1 min/month maximum drift						
Display	Color graphic LCD. Visible area 43 x 58 mm.						
USB Connection	1 USB port with mini USB connector.						
RS232C Connection	1 RS232C serial output with RJ12 (6P6C) connector for connection of a serial printer. Settable Baud Rate from 1200 to 115200.						
Auto-off	Configurable after 2, 5, 10, 15, 20 or 30 minutes from last pressure of a key, with battery supply. It can be disabled. Automatically disabled when powered from external supply.						
Operating conditions	-10 60 °C, 0 85% RH no condensation						
Storage temperature	-2565 ℃						
Materials	ABS - 55 shore rubber protection lateral bands and protective shell						
Dimensions	180 x 102 x 46 mm with rubber shell						
Weight	About 400 g. (with batteries and protection shell)						
Protection degree	IP 64 LCD description						
Status bar — 🛡 C							
A1~ A2	RH× 55.6 °C 24.79 – Measures						
Date / hour - 2015/05/	31.31 ¹⁸ Mon 08:31:16						

MENU

LOG

USB Functions



PRC Technologies Corp., Ltd. ลาดพร้ำว 101 กรุงเทพ 10240 โทรศัพท : 02 530 1714, 02 932 1711 มือถือ : 086 360 8600 อีเมล : contact@prctech.net LINE ID1 : prctec-info, LINE ID2 : @prctec

Probes and modules in line with the instrument

The probes of the instrument are equipped with an "intelligent" module acting as interface between the probe and the instrument. Inside the module there is a microprocessor circuit with permanent memory performing different functions:

- allows the instrument to recognize the type of connected probe
- stores into memory the probe calibration data
- recognizes the instrument with which it was calibrated
- maintains the factory-calibration and the last calibration data
- stores into memory a serial number allowing the unique identification of the probe.

TEMPERATURE PROBES

Temperature with platinum sensors (PRT)

- 4-wire Pt100 sensor temperature probes with SICRAM module

 For probes without SICRAM module, following modules are available: TP471 "intelligent" module with microprocessor and permanent memory, designed for 4-wire Pt100 sensors

TP47 connector without microprocessor and memory, designed for 4-wire Pt100 and 2 or 4-wire Pt1000 sensors.

Thermocouple probes

K, J, T, E, R, S, B and N-type thermocouple temperature probes can be connected. A SICRAM module has to be placed between the probe and the instrument:

TP471D0: 1-input module without cold junction compensation.

TP471D: 1-input module with internal sensor for cold junction compensation.

TP471D1: 2-input module with internal sensor for cold junction compensation

RELATIVE HUMIDITY AND TEMPERATURE COMBINED PROBES

The relative humidity and temperature combined probes of the HP47... and HP48... series use a capacitive sensor for the measurement of relative humidity and a Pt100 sensor for the measurement of temperature. By connecting these probes, the instrument can also display the following derived humidity quantities: **Saturated vapor pressure** (Svp) in hPa; **Partial vapor pressure** (Pvp) in hPa; **Mixing ratio** in g/kg (vapor grams in 1 kg of dry air); **Enthalpy** in J/g; **Absolute humidity** in g/m³ (vapor grams in 1 cubic meter of dry air); **Dew point Temperature** (Td); **Wet bulb Temperature** (Tw); **Discomfort index** (DiscIdx); **NET index** (NetIdx).



PRESSURE PROBES

For the measurement of pressure, following possibilities are available:

- PP471: SICRAM module for the measurement of absolute, relative and differential pressures. Operates with TP704... and TP705... series pressure probes.

- PP472: module for the detection of the barometric pressure in the range 600...1100 hPa.

- **PP4735...:** series of modules for the measurement of the **differential pressure.** Available full scale values: 2.5 (with auto-zero circuit) / 10 / 20 / 50 / 100 / 200 / 500 / 1000 / 2000 mbar.

AIR SPEED PROBES

Following probes can be connected to the instrument:

- AP471S...: hot-wire air speed SICRAM probes.
- AP472S...: vane air speed SICRAM probes.
- AP4735...: SICRAM modules for Pitot tubes.

PHOTOMETRIC AND RADIOMETRIC PROBES

The photometric and radiometric probes of **LP471... series** can be connected to the instrument. According to the probe used, the instrument can measure:

- illuminance,
- luminance,
- PAR (photon flux in the PAR chlorophyll field),
- irradiance, UVA UVB UVC and UVBC irradiance,
- effective irradiance in blue light spectral band,
- the combination of illuminance and of UVA irradiance,
- the weighted effective total irradiance according to UV action curve,
- global solar irradiance.

The connection of **pyranometers and albedometers** is also possible and can be done through the **VP472 module**. When connecting a pyranometer, the instument displays the mV signal generated by the thermopile and the global radiation in W/m². When connecting an albedometer, the HD31 displays the mV signal generated by the thermopile detecting incident and reflected light, the net radiation in W/m² and the albedo.

CO, PROBES

HD31.B3 and HD31.B3-10 are the probes available for the CO_2 measurement in the the ranges 0...5,000 or 0...10,000 ppm. The probes use a double wavelength infrared sensor with NDIR technology (Non-Dispersive Infrared Technology). The use of a double filter and of a particular measuring technique ensures accurate and stable measurement for a long time.

DIRECT VOLTAGE AND CURRENT

VP473 SICRAM module measures the direct voltage applied at the input in the range -20 Vdc to +20 Vdc. The module input impedance is 1 M Ω .

IP472 SICRAM module measures the **direct current** applied at the input in the range 0...24 mA. The module input impedance is 25Ω .





DeltaLog9 software

Deltalog9 application software can be easily download through the Delta OHM website. Once installed, it allows to manage the instrument simply and intuitively from your PC.

- Start and stop recording
- Direct download from PC of all the data contained in the instrument's memory card
- Parameters setting
- Data display as a table or a graph
- Real time display of the measurement data detected by the instrument and possibility to save, print and export in different format
- Reading of battery's charge status

Reading the measurements on the PC in real time (Monitor)

The instrument Monitor function displays the measurements registered by the instrument in real time on the PC's monitor, both graphically and numerically. The data input into the instrument are read and displayed according to a set update time from 1 second to 60 minutes. The data can be sent, according to set intervals, to an email address. The data received using the Monitor function are analysed in real time. Two alert limits can be associated to each measurement variable: an upper limit (High Level) and a lower limit (Low Level).





The HD31 is supplied with a strong and protective rubber housing, provided with a removable back support and a magnet.



In order to ensure the quality of our instruments, we are constantly re-evaluating our products. Improvements can imply changes in specification; we advise you to always check our website for the newest version of our documentation. PRC Technologies Corp., Ltd. ลาดพร้าว 101 กรุงเทพ 10240



DATA REPORT 00001 Monday, 2014/10/05 09:51:02							Model HD31xx Certified Multimeter				
	Date	Time	At some and	42 Temperature	Ad the r Painty Stry		Date	Time	Ar Recountly B	A2 Temperature	A2 Set V Persone *C Sty ho
00081	2014/20/06	12.46.02	\$2.1	:25-5	10 32.	08 10059	2014/10/06	29047157	42	6 25.	32 32.3
00002	2014/19/06	111-46-137	45.8	25.2	10 3.2.	08 235573	2014/10/25	08:47:18	43	6 25.	31 32.3
00003	2034/39/06	09-44-11	43.0	25.5	10 12.	11 CCC 91	2814/10/96	19147:19	43	6 25.	.33 32.5
00004	2014/10/06	09-44-12	\$3.8	36.3	10 12.	100 10077	0014/10/06	09:47:30	41	6 25.	.31 32.5
00005	2034/15/06	29.44-11	43.6	25.1	9 32	06 30071	3014/10/26	D8+47+31	43	6 25	31 32 5
00006	2034/39/04	00.44.18	63.7	25.0	9 32.	06 20074	2014/10/06	09+67+22	45.	6 25.	.31 32.3
00007	2014/19/06	09-46-15	13.7	25.3	9 32.	06 00075	2054/10/05	09.47:23	43	6 25	.31 32.3
00005	2014/19/06	65.46.12	43.7	25.7	9 52.	04 10071	2014/10/05	00.47.24	45.	6 25.	30 32.2
00005	2014/10/04	09-44-17	43.7	25.3	0 32.	08 00073	2014/10/36	09+47+25	43.	6 25.	.30 .32.2
00010	2014/10/04	00.44.18	42.7	25.1	0 32.	08 20078	2014/10/06	01147126	43.	6 25	30 32.2
00011	2014/16/04	09-45-19	\$3.7	29.3	.9 32.	06 30071	12014/10/06	09:47:27	43.	6 25.	30 32.2
00012	2014/10/04	03:44.28	43.7	25.3	9 52,	06 000 23	2014/10/35	00147128	45	6 25.	30 32.2
00013	2014/20/08	01-48-21	43.7	25.	9 32.	04 000 43	2014/10/38	199147123	43	8 25.	.32 32.3
00614	2014/15/06	09-46-33	\$1.1	26-3	H 57.	04 00083	3014/10/05	09+47-30	i 45.	6 25	40 82.2
00016	2014/10/04	00-46-23	43.6	31.3	.0 13.	04 00033	2014/10/06	09+47+31	45	6 25	31 32.3
00014	2014/10/04	00-46-28	0.1	35.3	0 33.	06 00038	2011/10/06	09+47+33	40	6 25	.01 02.5
00017	2014/10/04	00/48/28	45.4	25-2	.9 13.	06 000#1	2214/10/05	D9+47+38	0	8 25.	30 33.2
00010	2014/10/08	00.48.26	43.7	25.3	0 0.00	04209035	2014/10/05	09,47,34	40	6 25	2.20 00.2
00015	2014/10/04	05:44.27	43.7	25.3	.6 52.	04 200031	2064710/06	09,47,35	2 4 01	6 25.	.31 32.5
00020	2014/10/06	09-44-28	43.7	25.3	9 32.	06 000099	2814/12/26	09+47+34	43.	6 25	31 32.5
00021	2014/10/06	09/44/29	43.7	-25.3	9 12.	06 20033	2014/10/05	09147137	43	6 25.	32 32.5
00022	2814/19/04	10.441.00	43.7	25.1	9 52.	06 00090	2014/10/05	09147138	43.	6 25.	.32 32.5
00021	2014/10/08	0.0.44.11	6.1.5	36.11	10 A.C.	1100 200243	2014/13/34	04147130	43	6 25.	6.5k 54.
00024	2014/19/04	09-44-72	43.3	25.3	10 32	0# 00077	2014/10/06	05+47+40	41	6 25.	30 32.3
00025	2018/39/06	00 46 . 13	43.1	28.0		06 00073	2014/10/06	09+47+44	43.	# 25.	33 32.3
00026	2014/10/06	00.44.34	13.8	28-3	9 32.	06 30054	2014/10/04	08.47.43	44	4 25	34 32.3
00027	2014/10/00	00.46.25	63.9	25.3	9 32.	06 10001	2014/10/05	02.47.43	51	1 25.	36 32.5
00018	2034/30/06	00.46.34	43.10	25.1	10 32.	00 30030	2014/10/00	05,47,44	51	1 25	.40 32.4
00023	2014/30/06	00.46.37	12.1	25.3	10 33.	08 00097	2014/10/05	09147145	5 51.	1 25.	.47 32.6
00010	2014/10/06	12:44.38	13.11	25.2	11 32.	10 00098	2014/10706	01147146	49.	7 25.	.51 32.6
00001	2014/10/06	09:45:18	45.6	25.2	12 52.	22 30099	2014/10/05	09147147	49.	6 25	.53 32.7
000.2	2014/10/08	02:48.40	45.E	25.2	4 327	16 33172	2014/10735	09:47:48	49.	1 25	.53 32.7
00033	2014;13/06	09:44-41	45.8	36.3	M 52.	24回00101	8814/13/36	04:47:44	42,	7) 2h	54 3213
10011	100 A 8 8 18 A 19 A	D.D		76.1	10. 5.2	THE TRANSPORT	THE R & CR IS A MAN	nu. 19-14	2 414	T 95	E.70

Examples of reports in PDF format with graphs and tables automatically created by the instrument on the memory card.

Ordering codes

HD31 is supplied with lithium ion rechargeable battery, SD card, rubber protection shell with magnet, CP31 USB cable, SWD05 power supply and carrying case.

DeltaLog9 software is downloadable from Delta OHM website.

Modules, probes and serial cable for the printer have to be ordered separately.

CP31	USB cable with male mini-USB connector on instrument side and male A type USB connector on PC side.
CP31RS	RS232C connection cable for serial printer connection. RJ12 connector on the instrument side and Sub-D 9-pole female connector on the printer side.
SWD05	100-240 Vac / 5 Vdc-1 A stabilized mains power supply. Output with A type USB connector.
HD35-BAT1	Spare lithium ion rechargeable battery, 3.7 V, capacity 2250 mA/h, 3-pole JST connector.
HD40.1	Kit including 24-column portable thermal printer, serial interface, paper width 57mm, NiMH rechargeable battery pack, SWD10 power supply unit, 5 thermal paper rolls and instruction manual.
BAG31	Spare carrying case.

We look forward to your enquiry: Phone - +39 049 897 7150 Email - sales@deltaohm.com

Delta OHM S.r.l.

Single Member Company subject to direction and coordination of GHM MESSTECHNIK GmbH

Via Marconi 5 | 35030 Caselle di Selvazzano (PD) | ITALY

โทรศัพท : 02 530 1714, 02 932 1711 มือถือ : 086 360 8600 อีเมล : contact@prctech.net LINE ID1 : prctec-info, LINE ID2 : @prctec